

If Any Lift Equipment is NOT Working,

- 1) Report the problem to Clinical Engineering at 681-2525
- 2) Inform Nurse Manager & colleagues
- 3) Tag the lift "Out of Order"