

Tip of the Month: Proper Hovermatt Placement

Hovermatt air-assist device works best when positioned as so:

Green side down/
Gray side up
contacting patient



Foot End: Legs supported on
mat at end with inflator hose
opening & FOOT END label

Head End:

- Patient's head supported on mat
- Top of mat at top of patient's head

